

Development of Strip Readout Parallel Plate Avalanche Counter

Saturday, 24 August 2019 18:05 (15 minutes)

In a nuclear physics study using high intensity RI beam, high resolution position and timing detector is required for particle discrimination. We are developing Parallel Plate Avalanche Counter directly and independently reading out signals from strip electrodes, Strip Readout PPAC(SR-PPAC). It has no delay and high position resolution corresponding to strip size and charge resolution, and it is also stable even in high intensity beam without multi-hit. In this presentation, we report the evaluation of SR-PPAC through the results of

the previous experiments conducted in HIMAC in Chiba.

It is going to be implemented at RI Beam Factory at RIKEN towards the measurement of ^{132}Sn (p, p) for the study of the neutron skin.

Primary author: Mr HANAI, Shutaro (CNS)

Presenter: Mr HANAI, Shutaro (CNS)

Session Classification: Poster Session by Young Scientists